



**LABORATORY TESTING INC.**

2331 Topaz Drive, Hatfield, PA 19440  
Phone: 800-219-9095 ♦ Fax: 800-219-9096

*Certified Test Report*

SII006-07-03-06655-1



ISO/IEC 17025  
Materials Testing Laboratory  
NDT (PT, MT, UT)

**SOLD TO**

Sonotech, Incorporated  
774 Marine Drive  
Bellingham, WA 98225

**SHIP TO**

Sonotech, Incorporated  
774 Marine Drive  
Bellingham, WA 98225  
ATTN: Delilah Bragg

**CUSTOMER P.O.**

7724

**CERTIFICATION DATE**

3/12/2007

**SHIP VIA**

FAX AND MAIL

**DESCRIPTION**

Test Sample, ASTM D808, ASTM D129 and ASTM D512, 071-E, Covering: Echogel 30 - 07149, Echogel 20 - 07148; Echogel 14 - 07147; Echogel 10 - 07146, Including Winterized (all grades)

The submitted sample was analyzed in accordance with Customer's Instructions with the following results:

<u>ELEMENT</u>		<u>ACTUAL</u>
Total Halogens as Chloride (Cl, Br, I)	<	1. ppm
Total Sulfur		1. ppm

**Procedures/Methods:** MAS-ICP, Rev. 5, Analysis of Metals & Their Alloys by ICP

The services performed above were done in accordance with LTI's Quality System Program Manual Revision 17 dated 12/3/04 and ISO/IEC 17025. These results relate only to the items tested and this report shall not be reproduced, except in full, without the written approval of Laboratory Testing, Inc. L.T.I. is accredited by NADCAP to ISO/IEC 17025, Material's Testing and NDT (MT, PT and UT). Reported results represent the actual attributes of the material / item tested. Testing was performed in accordance with all applicable specification and contract / purchase order requirements.

**MERCURY CONTAMINATION:** During the testing and inspection, the product did not come in direct contact with mercury or any of its compounds nor with any mercury containing devices employing a single boundary of containment.

**NOTE:** The recording of false, fictitious or fraudulent statements or entries on this document may be punished as a felony under Federal Statutes.

Sherri L. Scheifele  
QA Specialist

By: \_\_\_\_\_  
Authorized Signature