



CHEMICAL TEST REPORT

Ref. C146720(N) Date December 7, 2009 Page 1 of 1

Customer: Sonotech, Inc., 774 Marine Drive, Bellingham, Washington 98225-1530
Attention: Judy Foley/Shari Brown

Purchase Order #: 112409 Part #/Name: 092-P

Material Designation: Couplant Material

Special Requirement: N/A

Lab Comment: Analyzed by ICP-atomic emission, ISE and titrimetric techniques.

Test Results

Composition: (parts per million)

Table with 7 columns: Identification, Total Cl, Total F, Total I, Total Br, Total Sulfur, Total Halogens. Rows include Alloy or Spec. Req., 092-P Covering, and various Pyrogel Gel samples.

- (1) Guidelines for analysis: ASTM D 808-05/ ASTM D 512-04
(2) Guidelines for analysis: ASTM D 808-05/ ASTM D 1179-04
(3) Guidelines for analysis: ASTM D 808-05 prep/ ICP-AE analysis
(4) Guidelines for analysis: ASTM D 129-00 prep/ ICP-AE analysis



Prepared by: [Signature] W. M. Katter Senior Chemist

Approved by: [Signature] D. M. McKay Supervisor

This report may not be reproduced except in full. This report represents interpretation of the results obtained from the test specimen and is not to be construed as a guaranty or warranty of the condition of the entire material lot. Measurement uncertainty is available upon request where applicable